



# CERTIFICATE OF ACCREDITATION

**The ANSI National Accreditation Board**

Hereby attests that

**DEWETRON Inc.  
2850 County Trail  
East Greenwich, RI 02818**

Fulfills the requirements of

**ISO/IEC 17025:2017**

In the field of

**CALIBRATION**

This certificate is valid only when accompanied by a current scope of accreditation document.  
The current scope of accreditation can be verified at [www.anab.org](http://www.anab.org).

Jason Stine, Vice President

Expiry Date: 02 June 2025

Certificate Number: AC-3212



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.  
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory  
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

## SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

**DEWETRON Inc.**  
2850 South County Trail  
East Greenwich, RI 02818  
401-284-3750

### CALIBRATION

Valid to: **June 2, 2025**

Certificate Number: **AC-3212**

#### Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Voltage – Source & Measure	Up to 100 mV (0.1 to 1) V (1 to 10) V (10 to 100) V (100 to 1 000) V	24 $\mu\text{V/V} + 0.3 \mu\text{V}$ 6.8 $\mu\text{V/V} + 0.3 \mu\text{V}$ 7.9 $\mu\text{V/V} + 0.5 \mu\text{V}$ 8.2 $\mu\text{V/V} + 30 \mu\text{V}$ 13 $\mu\text{V/V} + 0.1 \text{ mV}$	Fluke 5522A Multi Product Calibrator with Agilent 3458A Multimeter
DC Current – Source & Measure	Up to 1 mA (1 to 10) mA (10 to 100) mA (0.1 to 1) A (1 to 11) A (11 to 20) A	16 $\mu\text{A/A} + 5 \text{ nA}$ 27 $\mu\text{A/A} + 50 \text{ nA}$ 36 $\mu\text{A/A} + 0.5 \mu\text{A}$ 110 $\mu\text{A/A} + 10 \mu\text{A}$ 400 $\mu\text{A/A} + 0.39 \text{ mA}$ 770 $\mu\text{A/A} + 0.58 \text{ mA}$	Fluke 5522A Multi Product Calibrator with Agilent 3458A Multimeter
DC Current – Source & Measure	(0.2 to 1) A (1 to 10) A (10 to 20) A	0.003 2 % of reading + 2.3 $\mu\text{A}$ 0.004 5 % of reading - 11 $\mu\text{A}$ 0.007 6 % of reading - 320 $\mu\text{A}$	Agilent 3458A Multimeter with Fluke A40 Shunt
DC Resistance – Measure	Up to 10 $\Omega$ (10 to 100) $\Omega$ (0.1 to 1) k $\Omega$ (1 to 10) k $\Omega$ (10 to 100) k $\Omega$ (0.1 to 1) M $\Omega$	16 $\mu\Omega/\Omega + 50 \mu\Omega$ 8.0 $\mu\Omega/\Omega + 0.5 \text{ m}\Omega$ 6.9 $\mu\Omega/\Omega + 0.5 \Omega$ 7.4 $\mu\Omega/\Omega + 5 \Omega$ 7.2 $\mu\Omega/\Omega + 50 \Omega$ 12 $\mu\Omega/\Omega + 2 \Omega$	Agilent 3458A Multimeter



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Resistance - Source	Up to 11 $\Omega$	55 $\mu\Omega/\Omega$ + 1.0 m $\Omega$	Fluke 5522A Multi Product Calibrator
	(11 to 33) $\Omega$	47 $\mu\Omega/\Omega$ + 1.5 m $\Omega$	
	(33 to 110) $\Omega$	25 $\mu\Omega/\Omega$ + 1.4 m $\Omega$	
	(110 to 330) $\Omega$	22 $\mu\Omega/\Omega$ + 2.0 m $\Omega$	
	(0.33 to 1.1) k $\Omega$	29 $\mu\Omega/\Omega$ + 2.0 m $\Omega$	
	(1.1 to 3.3) k $\Omega$	22 $\mu\Omega/\Omega$ + 20 m $\Omega$	
	(3.3 to 11) k $\Omega$	25 $\mu\Omega/\Omega$ + 20 m $\Omega$	
	(11 to 33) k $\Omega$	37 $\mu\Omega/\Omega$ + 0.2 $\Omega$	
	(33 to 110) k $\Omega$	25 $\mu\Omega/\Omega$ + 0.2 $\Omega$	
	(110 to 330) k $\Omega$	27 $\mu\Omega/\Omega$ + 2 $\Omega$	
	(0.33 to 1.1) M $\Omega$	29 $\mu\Omega/\Omega$ + 2 $\Omega$	
	(1.1 to 3.3) M $\Omega$	73 $\mu\Omega/\Omega$ + 30 $\Omega$	
	(3.3 to 11) M $\Omega$	0.013 % of reading + 50 $\Omega$	
	(11 to 33) M $\Omega$	0.038 % of reading + 2.5 k $\Omega$	
	(33 to 110) M $\Omega$	0.061 % of reading + 3 k $\Omega$	
(110 to 330) M $\Omega$	0.4 % of reading + 0.1 M $\Omega$		
(330 to 1 100) M $\Omega$	1.2 % of reading + 0.5 M $\Omega$		
AC Voltage – Source & Measure	Up to 100 mV		Fluke 5522A Multi Product Calibrator with Agilent 3458A Multimeter
	(20 to 40) Hz	0.006 % of reading + 4 $\mu$ V	
	40 Hz to 20 kHz	0.015 % of reading + 2 $\mu$ V	
	(20 to 100) kHz	0.068 % of reading + 2 $\mu$ V	
	(0.1 to 1) V		
	(20 to 40) Hz	0.006 % of reading + 40 $\mu$ V	
	40 Hz to 20 kHz	0.014 % of reading + 20 $\mu$ V	
	(20 to 100) kHz	0.074 % of reading + 20 $\mu$ V	
	(1 to 10) V		
	(20 to 40) Hz	0.006 % of reading + 0.4 mV	
	40 Hz to 20 kHz	0.014 % of reading + 0.2 mV	
	(20 to 100) kHz	0.084 % of reading + 0.2 mV	
	(10 to 100) V		
	(20 to 40) Hz	0.016 % of reading + 4 mV	
	40 Hz to 20 kHz	0.017 % of reading + 2 mV	
(20 to 100) kHz	0.11 % of reading + 2 mV		
(100 to 1 000) V			
(20 to 40) Hz	0.043 % of reading + 40 mV		
40 Hz to 20 kHz	0.059 % of reading + 20 mV		
(20 to 100) kHz	0.3 % of reading + 20 mV		

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Source & Measure	(1 to 10) V (16 to 850) Hz	0.052 % of reading + 0.17 mV	Fluke 6105A Electrical Power Standard
	(10 to 100) V (16 to 850) Hz	0.05 % of reading + 1.3 mV	
	(100 to 1 000) V (16 to 850) Hz	0.05 % of reading + 9 mV	
AC Current – Source & Measure	Up to 100 µA 45 Hz to 5 kHz	0.06 % of reading + 30 nA	Fluke 5522A Multi Product Calibrator with Agilent 3458A Multimeter
	(0.1 to 1) mA 45 Hz to 5 kHz	0.06 % of reading + 0.2 µA	
	(1 to 10) mA 45 Hz to 5 kHz	0.06 % of reading + 2 µA	
	(10 to 100) mA 45 Hz to 5 kHz	0.06 % of reading + 20 µA	
	(0.1 to 1) A 45 Hz to 5 kHz	0.1 % of reading + 0.2 mA	
	(1 to 3) A 45 Hz to 5 kHz	0.48 % of reading + 0.78 mA	
	(3 to 11) A 45 Hz to 5 kHz	2.4 % of reading + 1.6 mA	
	(11 to 20) A 45 Hz to 5 kHz	2.3 % of reading + 3.9 mA	
	AC Current – Source & Measure	(0.01 to 0.25) A (16 to 850) Hz	
(0.25 to 1) A (16 to 850) Hz		0.006 % of reading + 20 µA	
(1 to 2) A (16 to 850) Hz		0.006 % of reading + 40 µA	
(2 to 5) A (16 to 850) Hz		0.006 4 % of reading + 100 µA	
(5 to 10) A (16 to 850) Hz		0.006 5 % of reading + 200 µA	
(10 to 21) A (16 to 850) Hz		0.007 1 % of reading + 400 µA	



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of Thermocouples - Measure/Source	Type J (-200 to -100) °C	0.3 °C	Fluke 5522A Multi Product Calibrator
	(-100 to 760) °C	0.2 °C	
	(760 to 1 200) °C	0.23 °C	
	Type K (-200 to -100) °C	0.35 °C	
	(-100 to 1 000) °C	0.25 °C	
	(1 000 to 1 370) °C	0.35 °C	
Electrical Simulation of RTDs - Measure/Source	Type T (-250 to -150) °C	0.64 °C	Fluke 5522A Multi Product Calibrator
	(-150 to 400) °C	0.17 °C	
	Pt385, 100 Ohm (-200 to 630) °C	0.11 °C	
	Pt3926, 100 Ohm (-200 to 630) °C	0.11 °C	
	Pt385, 200 Ohm (-200 to 630) °C	0.14 °C	
	Pt385, 500 Ohm (-200 to 600) °C	0.094 °C	
DC Power Source	Pt385, 1 000 Ohm (-200 to 600) °C	0.063 °C	Fluke 5522A Multi Product Calibrator with Agilent 3458A Multimeter
	10.9 μW to 1W	190 μW/W	
DC Power Source	(0.15 to 180) W	260 μW/W	Fluke 6105A Electrical Power Standard
	(180 to 720) W	260 μW/W	
	(720 to 2 016) W	260 μW/W	
	Up to 1 800 W	260 μW/W	
	(1 800 to 7 200) W	260 μW/W	
	(7 200 to 20 160) W	260 μW/W	
AC Power Source Power Factor PF = 1	16 Hz to 850 Hz (0.15 to 180) W	120 μW/W	Fluke 6105A Electrical Power Standard
	(180 to 720) W	120 μW/W	
	(720 to 2 016) W	110 μW/W	
	16 Hz to 850 Hz (1.5 to 1 800) W	130 μW/W	
	(1 800 to 7 200) W	120 μW/W	
	(7 200 to 20 160) W	120 μW/W	

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Power Source			
Power Factor PF = (< 1 to ≥ 0.9)	16 Hz to 450 Hz (0.15 to 20 160) W	200 μW/W	Fluke 6105A Electrical Power Standard
PF = (< 0.9 to ≥ 0.5)	16 Hz to 180 Hz (0.15 to 20 160) W	580 μW/W	
PF = (< 0.5 to ≥ 0.1)	45 Hz to 65 Hz (0.15 to 20 160) W	3 200 μW/W	

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ( $k=2$ ), corresponding to a confidence level of approximately 95%.

Notes:

1. This scope is formatted as part of a single document including Certificate of Accreditation No. AC-3212.



Jason Stine, Vice President

